



6330 Hollister Avenue  
Santa Barbara, CA 93117  
Tel. (805) 681-9009  
Fax (805) 681-0123  
email: info@wyatt.com  
www.wyatt.com

# Certificate of Analysis

## Sodium Chloride Solutions Kit

Sodium chloride (NaCl) in water:  $dn/dc = 0.174$  mL/g at 658 nm;  $dn/dc = 0.172$  mL/g at 690 nm

Wyatt p/n: P8400

Lot Number: 06042020

Expires: 6/4/2023

Storage: Room temperature

### Standard analysis data:

Standard	CAS#	Purity	Concentration (mg/mL)
18 MΩ water	7732-18-5	100.00 %	NEAT
0.100 mg/mL NaCl Solution	7764-18-5	99.999 %	0.100
0.250 mg/mL NaCl Solution	7764-18-5	99.999 %	0.250
0.500 mg/mL NaCl Solution	7764-18-5	99.999 %	0.500
1.000 mg/mL NaCl Solution	7764-18-5	99.999 %	1.000
1.200 mg/mL NaCl Solution	7764-18-5	99.999 %	1.200
2.000 mg/mL NaCl Solution	7764-18-5	99.999 %	2.000
3.000 mg/mL NaCl Solution	7764-18-5	99.999 %	3.000
4.000 mg/mL NaCl Solution	7764-18-5	99.999 %	4.000
5.000 mg/mL NaCl Solution	7764-18-5	99.999 %	5.000

Certification: Sophia Kenrick  
Sophia Kenrick  
Director of Analytical Services

Date: 6/18/2020